NELLCOR

Addendum to Patient Cable Instructions

Introduction

This addendum provides supplemental information to the *Directions for Use* for all applicable Nellcor manufactured or distributed pulse oximetry cables.

The information provided is divided into the following major topics:

- Applicability
- Safety
- Disposal
- Cables
- Compliance

Applicability

This addendum applies to all patient cables (listed in the Oximetry Cables table on page 2), unless noted otherwise.

Safety

Warning

The operator is responsible for checking the compatibility of the pulse oximetry monitor, sensor, and patient cable prior to use. Incompatible components can result in degraded performance and/or device malfunction.

Disposal

When retiring a cable from use, dispose of the device in accordance with all applicable laws and regulations.

10007890 Rev B (05/06)

Cables

The following table lists the oximetry cables intended for use with specific monitor model and OEM combinations.

Oximetry Cables

Nellcor Cable No.	Nellcor Monitors	Nellcor OEM Boards[1]
DEC-4 DEC-8	N-20 Series N-180 N-200 NPB-190 / NPB-195 NPB-4000 / NPB-4000 C NPB-3900 Series N-6000 *Ultra Cap*™ NPB-40 NPB-75™ *OxiMax*® NPB-75™ *OxiMax* N-85™	MP203 MP204 MP205 MP213 MP405 MP506 MP507 NELL-2 NELL-3
DEC-4 (Only)	*OxiMax* NPB-40 Max *OxiMax* N-65™	
SCP-10 MC-10	NPB-290 / NPB-295 N-395	MP304 MP404 MP405
SCP-10 (Only)	N-3000	
DOC-10	*OxiMax* N-550 *OxiMax* N-595 *OxiMax* N-600™ *OxiMax* N-560™	MP100 MP404 MP506 MP507 NELL-1 NELL-2 NELL-3

1. The individual OEM manufacturer determines whether Nellcor cables are used with specific monitors. If you have questions about which cable is to be used with which model of pulse oximeter, refer to the appropriate monitor's directions for use. In addition, the type of cable used depends upon how the board is incorporated into the OEM's monitor.

Compliance

When used with the appropriate Nellcor pulse oximetry monitors, patient monitors containing Nellcor oximetry, or with patient monitors licensed to use Nellcor oximetry, the pulse oximeter system is compliant with ISO 9919:2005, provided those system components meet all other requirements of this standard.

For more information, visit www.nellcor.com.